

Introduction

In a pulverized coal-fired boiler of a large power plant, an oxygen analyzer is essential for combustion control. Exhaust gases from the pulverized coal boiler contain a large quantity of dust and flow very fast. Oxygen analyzers that employ a sampling method may be subject to wear or clogging, resulting in increased maintenance workload and cost.

A solution to this problem is the ZR22G/ZR802G Zirconia Oxygen Analyzer that has no sampling system and utilizes a long-life sensor. A probe protector is attached to protection it against wear.

Expected Benefits

- Improves combustion efficiency for pulverized coal-fired boilers
- Ensures stable, continuous oxygen measurement
- Reduces operating costs
- Minimizes the need for equipment replacement

Process Overview

While heavy oil, light oil, kerosene commonly used as a boiler fuel, coal is also used because it is inexpensive and readily available. Unlike oil, coal and gas produce a large quantity of ashes when it is burned, necessitating pulverized coal-fired boilers to be equipped with an ash removal system such as a cyclone. Exhaust gases from these boilers contain a large quantity of dust (10 to 30 g/Nm³) and flow very fast as the result of the large volume of air being blown into the boiler. For oxygen measurement in large ducts, a probe with a long insertion length is used.



